61)

Notice of Allowability	Application No.	Applicant(s)
	10/797,714	NISHIMURA ET AL.
	Examiner	Art Unit
	John P. Leubecker	3739
The MAILING DATE of this communication appe All claims being allowable, PROSECUTION ON THE MERITS IS herewith (or previously mailed), a Notice of Allowance (PTOL-85) NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RI of the Office or upon petition by the applicant. See 37 CFR 1.313	(OR REMAINS) CLOSED in this app or other appropriate communication GHTS. This application is subject to	olication. If not included will be mailed in due course. THIS
1. This communication is responsive to <u>December 7, 2006</u> .		
2. ☑ The allowed claim(s) is/are <u>47-49</u> .		
 Acknowledgment is made of a claim for foreign priority una)	been received. been received in Application No	
Applicant has THREE MONTHS FROM THE "MAILING DATE" noted below. Failure to timely comply will result in ABANDONM THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.	of this communication to file a reply (ENT of this application.	complying with the requirements
4. A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.		
5. CORRECTED DRAWINGS (as "replacement sheets") must be submitted.		
(a) 🔲 including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached		
1) 🗌 hereto or 2) 🔲 to Paper No./Mail Date		
(b) ☐ including changes required by the attached Examiner's Paper No./Mail Date	s Amendment / Comment or in the O	office action of
Identifying indicia such as the application number (see 37 CFR 1 each sheet. Replacement sheet(s) should be labeled as such in the		
DEPOSIT OF and/or INFORMATION about the deposit attached Examiner's comment regarding REQUIREMENT I		
Attachment(s) 1. ☑ Notice of References Cited (PTO-892) 2. ☑ Notice of Draftperson's Patent Drawing Review (PTO-948) 3. ☑ Information Disclosure Statements (PTO/SB/08),	5. ☐ Notice of Informal P 6. ☐ Interview Summary Paper No./Mail Dat 7. ⊠ Examiner's Amendn	(PTO-413), e
Paper No./Mail Date 3/10/04 4. Examiner's Comment Regarding Requirement for Deposit of Biological Material	_	ent of Reasons for Allowance

Application/Control Number: 10/797,714

Art Unit: 3739

EXAMINER'S AMENDMENT

1. An examiner's amendment to the record appears below. Should the changes and/or

additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR

1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the

payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with

Thomas Spinelli on March 1, 2007.

The application has been amended as follows:

IN THE ABSTRACT

Replace the Abstract with the following:

--An endoscope inserting direction detecting system receives an endoscopic image,

detects an endoscope inserting direction, in which an endoscope should be inserted, on the basis

of the endoscopic image and produces information concerning the detected inserting direction.

A plurality of detecting algorithms, each of which detect the inserting direction, are included and

any of the plurality of detecting algorithms can be selected based on the endoscopic image.—

IN THE CLAIMS

Cancel claims 1-46 and 50-53.

In claim 49, line 1, replace "method" with -system--.

Page 2

Art Unit: 3739

2. The following is an examiner's statement of reasons for allowance: This application is in condition for allowance except for the presence of claims directed to an invention non-elected without traverse. Accordingly, claims 1-46 and 50-53 have been cancelled. It is noted that claim 47 was erroneously included in the non-elected claims. That claim has been examined with claims 48 and 49.

All references of record show endoscope inserting direction detecting systems and methods (e.g., Gillies et al. (4,910,590) and Suzuki et al. (5,018,509)) that are based on the endoscope image. Although each reference utilizes a plurality of computations to obtain a insertion direction result, all computations are part of one detecting algorithm. Thus, the prior art of record fail to teach or suggest the combination of claim 48 including a detecting algorithm changing means for changing detecting algorithms, any of which the inserting direction detecting means adopts, according to the endoscopic image. Method claim 47 recites, *inter alia*, that a plurality of detecting algorithms are included in order to detect the inserting direction and any of such algorithms is selected based on the endoscope image. Claim 47 is thus allowable for the same reason as set forth above.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to John P. Leubecker whose telephone number is (571) 272-4769. The examiner can normally be reached on Monday through Friday, 6:00 AM to 2:30 PM.

Art Unit: 3739

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda C.M. Dvorak can be reached on (571) 272-4764. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

John P. Leubecker Primary Examiner Art Unit 3739

jpl